

THE EPPLEY LABORATORY, INC.

12 Sheffield Avenue, PO Box 419, Newport, Rhode Island USA 02840 Phone: 401.847.1020 Fax: 401.847.1031 Email: info@eppleylab.com

NORMAL INCIDENCE PYRHELIOMETER MODEL SNIP



A pyrheliometer mounted on a solar tracker is used to measure the Direct Beam Solar Irradiance (DNI) from the sun. Historically, the preferred field of view for Pyrheliometers was based on a 10:1 ratio which equated to approximately 5.7°. Due in part to the commercialization of the Eppley AHF Cavity Radiometer as a Primary Standard and advances in accuracy of Automatic Solar Trackers (such as the Eppley SMT Tracker), the preferred FOV for pyrheliometers is now 5° which the Eppley sNIP uses. In fact, the sNIP has the exact same geometric dimensions as used in the AHF. Compared to the older NIP, the sNIP also has a faster response time, reduced conduction and convection effects and a thermistor is included for those who wish to measure the instrument temperature.

As a result, the Normal Incidence Pyrheliometer, Model sNIP meets the performance specifications of an ISO Secondary Standard* and a WMO High Quality Pyrheliometer

MODEL sNIP SPECIFCATIONS

Application: Standard/Network Measurements
Classification: Secondary Standard*/High Quality
World Radiation Reference (WRR)

Spectral Range 250-3000 nm Field of View 5°

Output 0-10 mV

Sensitivity approx. $8 \mu V / Wm^{-2}$ Impedance approx. 200Ω

95% Response Time
Zero Offset
Non-Stability
Non-Linearity
Spectral Selectivity
Temperature Response

5 seconds
1 Wm⁻²
0.5%
0.5%

Calibration Uncertainty** < 1% Measurement Uncertainty**

Single Point < 5 Wm⁻²
Hourly Average approx. 1%
Daily Average approx. 1%

Since the dawn of time, man has studied the sun...
...and Eppley has been providing the best instruments since 1917!

^{*} To officially be considered a Secondary Standard, the pyrheliometer in question must be calibrated with WRR traceability through a Primary Standard Pyrheliometer such as the Eppley AHF Cavity Radiometer. EPLAB Calibrations are typically performed against a Secondary Standard Pyrheliometer. At the customer's request and for an additional fee, this calibration can be performed against our WRR traceable AHF Cavity Radiometer. Please contact Eppley for additional information.

^{***} There has been much discussion on "uncertainty" and how it pertains to solar measurements. The RSS of the 9060 specifications results in an uncertainty of approximately 1.5%. The typical uncertainty of Eppley's factory calibrations are less than 1%. The stated uncertainty of the WRR is 0.4%. Evidence from direct comparisons of sNIP to AHF show the sNIP is capable of hourly and daily averages better than 1% (assuming proper tracking and clean windows).